

JRA10:CryPTA

Cryogenic Polarized Target Applications

Hartmut Dutz
Physikalisches Institut Universität Bonn

This project has received funding from the European Union's Horizon 2020 research and innovation programme under grant agreement No 824093



JRA10:CryPTA

Cryogenic Polarized Target Applications

Cooperation of four partners



Organization legal name	Short name	Activity leaders
Ruder Boskovic Institute	RBI	M. Korolija
Ruhr-Universität Bochum	RUB	G. Reicherz
Rheinische Friedrich-Wilhelms-Universität Bonn	UBO	H. Dutz
Johannes Gutenberg Universität Mainz	UMainz	A. Thomas



Research Objectives

The final goal of CryPTA is to develop groundbreaking s.c. magnet structures and low temperature detector techniques for new and innovative polarization experiments using polarized targets in 4π -detection systems for hadron physics experiments in Europe

JRA10:CryPTA Cryogenic Polarized Target Applications


Task 1 CryPTA:SuperconductingMagnet for permanent DNP in polarized solid state targets.
Polarization experiments @ ELSA and MAMI (i.e. baryon spectroscopy)
→ Modell independent partial wave analysis, complete experiment

Task 2 CryPTA:SuperconductingShield for passive shielding of the PANDA spec. field for transv. polarization

High temperature ScM for passive or active magnetic field shielding for polarization experiments with internal pol. targets @ PANDA.

Imaginary part of time like FF single spin target asymmetry → transverse polarization


Task 3 CryPTA:ActivePolarizedTarget materials to detect the recoil proton in-situ, inside the polarized target @ MAMI: Measurements of the Proton Spin-Polarizabilities with Double-Polarized Compton Scattering



Main problems:


- Low energetic recoil protons do not escape from the target and do not reach the detector.
- Events are produced on the background nuclei (Carbon, coherent, incoherent, $k \sim 13\%$).

→ detect the recoil proton in-situ, inside a highly polarized target material




Photon polarization	Target polarization	Recoil nucleon polarization	Target and recoil polarizations
	X Y Z(beam)	X' Y' Z'	X' X' Z' Z' X Z X Z
unpolarized	σ - T -	- P -	T' L' T' L'
linear	Σ H (-P) -G	O' (-T) O'	(-L') (T') (L') (-T')
circular	F - -E	C' - C'	- - - -

Permanent high dyn. Polarization @ data taking



JRA10:CryPTA


Cryogenic Polarized Target Applications






TASKS/Subtasks	Year 1				Year 2				Q1
	Q1	Q2	Q3	Q4	Q1	Q2	Q3	Q4	
1. Development of low mass superconducting high field magnets									
1.1 High precision winding machine for thin superconducting wires			1						
1.2 Manufacture and test of a small size low mass polarizing solenoid with high homogeneity									
1.3 Design manufacture and cold test of a prototype low mass, combined field superconducting magnet system									2
1.4 Magnet field design studies for a low mass large sc. tracking solenoid									
2. Development of low mass superconducting passive shielding									
2.1 Magnet field calculations for PANDA low mass superconducting passive shielding									3
2.2 Design and Manufacture of prototype HTSC shields and test at cryogenic temperatures									
3. Detection of recoil particles in active polarized targets at cryogenic temperatures									
3.1 Design studies for polarized, scintillating target material									
3.2 Prototypes of a scintillating target stacks with electronic readout									4
3.3 Prototype of a new cryogenic insert with active target material									

JRA10:CryPTA:Task 1

1.2 Manufacture and test of a small size low mass polarizing solenoid with high homogeneity




- Small superconducting magnet designed and build for DNP (2.5 T) in a dilution refrigerator will be wound soon on the new winding machine.
 - N=5116, 8 Layers, 2.5T @ 68.6A/1K, thickness 2.4 mm
- New ⁴He evaporation refrigerator as simple 1K test facility for superconducting magnets is under construction
 - Fits into the external high field DNP-magnet ($B_{max} = 6.5T$)
 - Field measurement (mapping) by pulsed-, cw-NMR, Hall-probe
 - Equipped for DNP (50 – 140 GHz)
 - Variable temperature range (1K – 70K)
 - Flexible and open access via insert tube
 - Large low temperature volume for magnet tests (500 mm x \varnothing 75 mm, 2.2l)
- New magnet has to be tested under nominal conditions (1K).
- 1K test facility will be available first half next year






JRA10:CryPTA:Task 1

1.3 Design, manufacture and cold test of a prototype low mass combined field superconducting magnet system

Combined longitudinal and transverse holding coil for the new dilution refrigerator for a variable polarization direction in plane (Milestone MS64)



- Cu-support (cooling)
- Longitudinal field: solenoid (outside)
 - 4 layers á N590
 - $I_{max} \sim 32$ A
 - $B_{max} \sim 0.52$ T
- Transverse field: race track like dipole (inside)
 - 2 x N525
 - $I_{max} \sim 30$ A
 - $B_{max} \sim 0.50$ T




- Precision wet winding of a solenoid is a well established technique.
- Wet winding of race tracks with thin superconducting wires has to be improved to guaranty high performance at minimum thickness
- Problem: inhomogeneous glue distribution within the coil package
- New wire feeding device is under construction

JRA10:CryPTA:Task 1


1.3 Design, manufacture and cold test of a prototype low mass combined field superconducting magnet system

New horizontal dilution refrigerator for polarization experiments with Crystal Barrel detector @ ELSA
Build by cryogenic department of JINR (Dubna)



Test set-up in the PT-laboratory

Central beam line
Target insert for fast and easy target exchange
Designed for various magnet configurations
→ $T_{\min} < 30 \text{ mK}$, TDNP $\sim 250 \text{ mK}$
→ $I_{\max} \sim 40 \text{ A}$




- Delivered fall 2020, assembled and first 1K tests this summer, commissioning in spring 2022
- First polarization experiment @ ELSA planned for fall 2022 using the combined sc.coil configuration
- Foreseen for pol. experiments using active polarized target technology

JRA10:CryPTA:Task 3



3.1 Design studies for polarized, scintillating target material

New ideas for a readout with optical fibers have been investigated and further developed in the framework of the finished PhD work of M.Biroth. He has been engaged using EU funding in the period 4-9/2021 to proceed the conception of the system.



1st problem of the prototype:

The transverse integral thermal expansion of PMMA and stainless steel (SS) differs by a factor of six. This has to be compensated by a **relaxation pod, which takes the stress** from the tube-flange connection.




Spring construction

2nd problem of the prototype:

The longitudinal thermal expansion of ~1% of the 1.5 m PMMA tube has to be compensated at the detector board holder by springs.

JRA10:CryPTA:Task 3


3.2 Prototypes of a scintillating target stack with electronic readout



Prototypes for an readout with optical fibres have been investigated.

Advantages:


- Better light coupling and transport.
- A semi active design using Butanol with better dilution factor becomes possible. D-Butanol is another option.
- Thermal expansion can be partially compensated.



Flange and tube (SS)



Scintillating head with fiber readout



Butanol container (PTFE)

→ the preparation of an improved target insert with optical fiber readout is on the way

List of Milestones in the reporting period

Milestone number	Milestone name	Lead beneficiary	Delivery month from Annex I	Delivered (yes/no)	Actual delivery month	Comments
MS64	Design concept of a low mass, combined field superconducting magnet system	10 – UBO	27	yes	29	A detailed description of the milestone will be available on: https://www.polarisiertes-target.physik.uni-bonn.de/startseite/strong-2020/status-und-eckpunktberichte
MS65	Magnet field calculations for PANDA low mass superconducting passive shielding	10 – UBO	27	yes	27	
MS66	Manufacture of prototype active targets for in beam tests	10 – UBO	27	yes	27	

No Deliverables in the reporting period

- CryPTA annual meeting 2021, June 24th : <https://www.polarisiertes-target.physik.uni-bonn.de/startseite/strong-2020/status-und-eckpunktberichte/statusberichte/crypta-jahrestreffen-2021>
- CryPTA workshop planned for spring 2022 in Croatia (RBI)

Research Objectives

The final goal of CryPTA is to develop groundbreaking s.c. magnet structures and low temperature detector techniques for new and innovative polarization experiments using polarized targets in 4π -detection systems for hadron physics experiments in Europe

Despite the currently unfavorable circumstances: the 3 Tasks are (up to now) on the right track

- Task 1: → completing the small size low mass polarizing solenoid with high homogeneity and commissioning of the magnet test refrigerator.
 - preparing a combined magnet for the new dilution refrigerator and the near future (2022) data taking campaign.
- Task 2: → Magnet field calculations for PANDA low mass superconducting passive shielding has been successfully completed.
- Task 3: → detailed design concepts for low temperature polarized active targets are defined.
 - the preparation of an improved target insert with optical fiber readout is on the way .

But: Corona is affecting more and more the work and progress of JRA10

So: lets hope the best

More information about CryPTA: <https://www.polarisiertes-target.physik.uni-bonn.de/startseite/strong-2020/>