Eleventh International Workshop on Semiconductor Pixel Detectors for Particles and Imaging



ID de Contribution: 18

Type: 12mOral

100µPET: an ultra-high-resolution silicon-pixel-based PET scanner

mardi 19 novembre 2024 11:15 (17 minutes)

The 100 μ PET project is developing a pre-clinical medical scanner for positron-emission tomography (PET) with ultra-high-resolution molecular imaging capabilities. The scanner is composed of multiple layers of monolithic active pixel sensors (MAPS) connected to flexible printed circuits (FPC). With pixels of 150 μ m pitch and a thickness of 280 μ m + 300 μ m (MAPS + FPC), the scanner achieves unprecedented volumetric spatial resolution of 0.02 mm³, one order of magnitude better than the best current PET scanners, and offers uniform resolution along the scanner's field-of-view (parallax free). The MAPS and its design features will be presented, along with the pixel read-out architecture. The construction and quality control of the scanner and its multiple detection modules, prototyped with pre-production chips and FPCs, will be showcased, and the latest imaging reconstruction with simulated high-definition mouse phantoms will be presented.

Auteurs principaux: IACOBUCCI, giuseppe (Université de Genève); CARDELLA, roberto (niversité de Genève)

Orateur: CARDELLA, roberto (niversité de Genève)

Classification de Session: Medical imaging applications

Classification de thématique: Applications in biology, medical imaging