

Patch panel for m2 Micromegas chambers

- Motivation

Parallel distribution of gas to several m2 chambers (3 liters each) for future beam tests

- Requirements

- 1 gas line/chamber with gas and flow control
- Pure Ar for purging chambers
Several chamber volume / hour → Up to 15 l/h
- Premix Ar/iso

